

IMPORTANT: Please read carefully and keep
this information for future use.

This package insert is intended for the eyecare
practitioner, but should be made available to the
patient upon request. The eyecare practitioner
should provide the patient with the wearer's
guide that pertains to the patients prescribed
lens.

**eyedia® fresh daily disposable (HIOXIFILCON A) DAILY DISPOSABLES
SOFT CONTACT LENS (CLEAR AND VISIBILITY TINT)**

eyedia® **fresh** daily
disposable

***CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE TO SALE BY OR ON
THE ORDER OF A LICENSED PRACTITIONER.***

clearlab®

clearlab® is trademark of Clearlab SG Pte. Ltd. and eyedia® is trademark of Clearlab US, Inc.

DESCRIPTION:

The eyedia® fresh daily disposable (**Hioxifilcon A**) Soft Contact Lens is available as a single vision spherical lens. The hydrophilic nature of this material allows the lens to become soft and pliable when immersed in an aqueous solution.

The non-ionic lens material, (hioxifilcon A) is a random co-polymer of 2-hydroxyethyl methacrylate and glycerol methacrylate cross-linked with ethylene glycol dimethacrylate. It consists of 42% hioxifilcon A and 58% water by weight when immersed in a buffered saline solution. The lens is available with a pale blue visibility handling tint, color additive ‘Reactive Blue 4’ 21 CFR Part 73.3121. The hioxifilcon A name has been adopted by the United States Adopted Names Council (USAN).

In the hydrated state, the lens conforms to the curvature of the eye covering the cornea and extending slightly beyond the limbus forming a colorless, transparent optical surface. The hydrophilic properties of the lens require that it be maintained in a fully hydrated state in a solution compatible with the eye. If the lens dries out, it will become hard and appear somewhat warped however, it will return to its proper configuration when completely rehydrated in the proper storage solution.

The hydrophilic characteristics allows aqueous solutions to enter the lens and in it’s fully hydrated state and the lens is approximately 58% water by weight. The physical properties of the lens are:

Refractive Index at 21°C	1.4011(wet)
Light Transmission	> 95%
Surface Character	Hydrophilic
Water Content at 21°C	59.77%
Specific Gravity at 21°C	1.086(wet)
Oxygen Permeability at 34-36 °C	25.38 (cm ² /sec)(ml O ₂ /ml x mmHg), (revised FATT method)

The lenses are hemispherical flexible shells which cover the cornea and a portion of the adjacent sclera with the following dimensions:

Chord Diameter:	14.0-14.4 mm
Centre Thickness:	0.09 mm
Base Curve:	8.5-8.10 mm
Powers:	+6.00 to +0.50 Dioptres in 0.25 D increment -0.50 to -6.00 Dioptres in 0.25 D increment -6.50 to -10.00 Dioptres in 0.50 D increment

ACTIONS:

In its hydrated state, the **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens**, when placed on the cornea, acts as a refracting medium to focus light rays on the retina.

INDICATIONS (USES):

The **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** is indicated for daily wear single use only for the optical correction of refractive ametropia (myopia and hyperopia) in not-aphakic and aphakic persons with non-diseased eyes who may have 1.00D or less of astigmatism.

The lens is intended to be worn once and then discarded at the end of each wearing period on a daily basis. The patient should be instructed to start the next wearing period with a new lens.

CONTRAINDICATIONS (REASONS NOT TO USE):

DO NOT USE the **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** when any of the following conditions are present:

- Acute and subacute inflammation or infection of the anterior chamber of the eye.
- Any eye disease, injury, or abnormality that affects the cornea, conjunctiva, or eyelids.
- Severe insufficiency of lacrimal secretion (dry eyes).
- Corneal hypoesthesia (reduced corneal sensitivity), if not-aphakic.
- Any systemic disease that may affect the eye or be exaggerated by wearing contact lens.
- Allergic reactions of ocular surfaces or adnexa that may be induced or exaggerated by wearing contact lens or use of contact lens solutions.
- Any active corneal infection (bacterial, fungi, or viral).
- If eyes become red or irritated.
- Patients unable to follow lens care regimen or unable to obtain assistance to do so.
- Advise patient not to wear **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** while sleeping.

WARNINGS:

- **PROBLEMS WITH CONTACT LENS COULD RESULT IN SERIOUS INJURY TO THE EYE.** It is essential that you follow your eyecare practitioner's direction and all labeling instructions for proper use of lens and lens care products, including the lens case. **EYE PROBLEMS, INCLUDING CORNEAL ULCERS, CAN DEVELOP RAPIDLY AND LEAD TO LOSS OF VISION.**

- Daily wear lens are not indicated for overnight wear, and you are instructed not to wear the **EYEDIA® FRESH DAILY DISPOSABLE (Hioxifilcon A) Daily Disposables** Soft Contact Lens while sleeping. Clinical studies have shown that the risk of serious adverse reactions is increased when these lens are worn overnight.
- Studies have shown that contact lens wearers who are smokers have a higher incidence of adverse reactions than nonsmokers do.
- **IF YOU EXPERIENCE EYE DISCOMFORT, EXCESSIVE TEARING, VISION CHANGES, OR REDNESS OF THE EYE, IMMEDIATELY REMOVE YOUR LENS AND PROMPTLY CONTACT YOUR EYECARE PRACTITIONER.**

PRECAUTIONS:

Special Precautions for the Eyecare Practitioners:

- Clinical studies have demonstrated that the contact lenses manufactured from (hioxifilcon A) are safe and effective for their intended use. However, the clinical studies may not have included all the design configurations or lens parameters that are presently available in this lens material.
- Always discard disposable lenses after the recommended wearing schedule prescribed by the eyecare practitioner.
- Due to the small number of patients enrolled in clinical investigation of lenses, all refractive powers, design configurations, or lens parameters available in the lens material are not evaluated in significant numbers. Consequently when selecting an appropriate lens design and parameter, the eyecare practitioner should consider all characteristics of the lens that can affect lens performance and ocular health, including oxygen permeability, wettability, central and peripheral thickness and optic zone diameter.
- The potential impact of these factors on the patient's ocular health must be carefully weighed against the patient's need for refractive correction. Therefore the continuing ocular health of the patient and the lens performance on the eye should be carefully monitored by the prescribing eyecare practitioner.
- Aphakic patients should not fitted with the **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** until the determination is made that the eye has healed completely.
- Fluorescein, a yellow dye, should not be used while the lens is on the patient's eye. The lens absorbs this dye and becomes discolored. Whenever Fluorescein is used in eyes, the eyes should be flushed with a sterile saline solution that is

recommended for in eye use. Wait at least one hour before replacing the lens. Too early replacement may allow the lens to absorb residual Fluorescein irreversibly.

- Before leaving the eyecare practitioner's office, the patient should be able to promptly remove the lens or should have someone else available to remove the lens for him or her.
- Eyecare practitioners should instruct the patient to remove the lens immediately if the eye becomes red or irritated.

The eyecare practitioner should carefully instruct patients about the following lens care regimen and safety precautions:

- The lens is intended to be worn once and then discarded at the end of each wearing period.
- The patient should be instructed to start the next wearing period with a new lens.
- Emergency lens cleaning and disinfection is not recommended.
- The patient should be reminded to have replacement lenses or back-up spectacles available at all times.
- The patient should be instructed to always have a spare pair of lenses at all times.
- Always wash and rinse hands before handling lens.
- Do not get cosmetics, lotions, soaps, creams, deodorants, or sprays in the eyes or on the lens.
- It is best to put on lens before putting on makeup. Water-base cosmetics are less likely to damage lens than oil-base.
- Do not touch contact lens with the fingers or hands if the hands are not free of foreign materials, as microscopic scratches of the lens may occur, causing distorted vision and/or injury to the eye.
- Carefully follow the handling, insertion, removal and wearing instructions in the patient instructions for the **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** and those prescribed by the eyecare practitioner.
- Never wear lens beyond the period recommended by the eyecare practitioner.
- If aerosol products such as hair spray are used while wearing lens, exercise caution and keep eyes closed until the spray has settled.
- Always handle lens carefully and avoid dropping them.
- Avoid all harmful or irritating vapors and fumes while wearing lens.
- Ask the eyecare practitioner about wearing lens during sporting activities.
- Inform the doctor (health care practitioner) about being a contact lens wearer.
- Never use tweezers or other tools to remove lens from the lens container unless specifically indicated for that use. Pour the lens into the hand.
- Do not touch the lens with fingernails.
- Always contact the eyecare practitioner before using any medicine or medications in the eyes.

- Always inform the employer of being a contact lens wearer. Some jobs may require use of eye protection equipment or may require that the patient not wear contact lens.
- As with any contact lens, follow-up visits are necessary to assure the continuing health of the patient's eyes. The patient should be instructed as to a recommended follow-up schedule.

ADVERSE REACTIONS:

The patient should be informed that the following problems may occur:

- Eyes stinging, burning, itching (irritation), or other eye pain.
- Comfort is less than when lens was first placed on eye.
- Feeling that something is in the eye such as a foreign body or scratched area.
- Excessive watering (tearing) of the eye.
- Unusual eye secretions.
- Redness of the eye.
- Reduced sharpness of vision (poor visual acuity).
- Blurred vision, rainbows, or halos around objects.
- Sensitivity to light (photophobia).
- Dry eyes.

If the patient notices any of the above he or she should be instructed to:

- **IMMEDIATELY REMOVE THE LENS.**
- If discomfort or problems stops, then look closely at the lens. If the lens is in a way damaged, **DO NOT PUT THE LENS BACK ON YOUR EYE.** Discard the lens and replace with a fresh lens. Place the lens in the storage case and contact the eyecare practitioner. If the lens has dirt, an eyelash, or other foreign body on it, or the problem stops and the lens appears undamaged, the patient should discard the lens and replace with a new one. **After reinsertion, if the problem continues, you should IMMEDIATELY REMOVE THE LENS AND CONSULT YOUR EYECARE PRACTITIONER.**

When any of the above problems occur, a serious condition such as infection, corneal ulcer, neovascularization, or iritis may be present. **KEEP THE LENS OFF THE EYE AND SEEK IMMEDIATE PROFESSIONAL IDENTIFICATION** of the problem and prompt treatment to avoid serious eye damage.

FITTING:

Conventional methods of fitting contact lenses apply to the **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens**. For detailed description of the fitting techniques, refer to the **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens Professional Fitting Guide**, copies of which are available from:

PACKAGE INSERT



Clearlab US Inc.

4200 Jenkins Court, Suwanee,
GA 30024, United States of America
Tel: +1 770 2710211
Fax: +1 770 2710225
Email: USRA@clearlab.com
Website: www.clearlabus.com

WEARING SCHEDULE:

THE WEARING AND REPLACEMENT SCHEDULE SHOULD BE DETERMINED BY THE EYECARE PRACTITIONER.

Patients tend to overwear the lens initially. It is important not to exceed the wearing schedule. The eyecare practitioner should emphasize the importance of adhering to the initial maximum wearing schedule. Regular check ups, as determined by the eyecare practitioner, are also extremely important.

The **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** are indicated for daily wear. The maximum suggested wearing time for the **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** is:

DAY	1	2	3	4	5	6
HOURS	6	8	10	12	14	All waking hours

STUDIES HAVE NOT BEEN COMPLETED TO SHOW THAT THE eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens" IS SAFE TO WEAR DURING SLEEP.

WEARING SCHEDULES SHOULD BE DETERMINED BY THE EYECARE PRACTITIONER

PLANNED REPLACEMENT:

The **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** is intended to be worn once and then discarded at the end of each wearing period. The patient should be instructed to start the next wearing period with a new lens.

LENS CARE DIRECTIONS:

The **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** is intended to be worn once and then discarded at the end of each wearing period. The patient should be instructed to start the next wearing period with a new lens.

PACKAGE INSERT



Eyecare Practitioners should review with the patient lens care directions, including both basic lens care information and specific instruction on the lens care wearing schedule recommended for the patient.

EMERGENCIES:

The patient should be informed that if any chemical of any kind (household products, gardening solutions, laboratory chemicals, etc) are splashed into the eyes, the patient should:

FLUSH EYES IMMEDIATELY WITH TAP WATER AND IMMEDIATELY CONTACT YOUR EYECARE PRACTITIONER OR VISIT A HOSPITAL EMERGENCY ROOM WITHOUT DELAY.

HOW SUPPLIED:

Each lens is supplied sterile in blister packs containing buffered saline solution. The blister pack is marked with dioptric power, Single Patient Use, Rx symbol, sterile symbol, composition of the lens, manufacturing lot number and expiration date of the lens. The blister pack is also marked as 'NOT FOR INDIVIDUAL RESALE'.

REPORTING OF ADVERSE REACTION:

Practitioners should report any adverse reactions to **eyedia® fresh daily disposable (Hioxifilcon A) Soft Contact Lens** within 5 days to at the address below:

Additional Professional Fitting Guide and Patient Instruction/ Wearer's Guide are available from:

Clearlab US Inc.

4200 Jenkins Court, Suwanee,
GA 30024, United States of America

Tel: +1 770 2710211





Fax: +1 770 2710225

Email: USRA@clearlab.com

Website: www.clearlabus.com

SYMBOLS KEY

The following symbols may appear on the label or carton.

SYMBOL	DESCRIPTION
	Sterile Using Steam
LOT	Product Lot Number
EXP	Expiry Date
BC	Lens Base Curve
DIA	Lens Diameter
	Do not reuse
Rx Only	Caution: Federal law restricts this device to sale by or on the order of a licensed Eye Care Practitioner.
 SINGLE PATIENT USE	Caution: See Instruction Leaflet; this is a single patient use device.
	Packaging material is recyclable

clearlab®

Clearlab US Inc.
 Printed in Singapore
 Revision Date: 05 October 2011
 Doc Number: S-ASP-017-A
 Version Number: V02

© 2011 CLEARLAB SG PTE. LTD. All Rights Reserved.