

IMPORTANT: Please read carefully and keep this information for future use.

This package insert is intended for the eyecare practitioner, but should be made available to the patient upon request. The eyecare practitioner should provide the patient with the wearer's guide that pertains to the patients prescribed lens.

eyedia® **precise** monthly for astigmatism (HIOXIFILCON A)
DAILY WEAR SOFT CONTACT LENS

eyedia® **precise** monthly for astigmatism

CAUTION: FEDERAL LAW RESTRICTS THIS DEVICE TO SALE BY OR ON THE ORDER OF A LICENSED PRACTITIONER.

clear**lab**®

clearlab® is trademark of Clearlab SG Pte. Ltd. and eyedia® is trademark of Clearlab US, Inc.

IMPORTANT:

Please read carefully and keep this information for future use. This package insert is intended for the eyecare practitioner, but should be made available to the patient upon request. The eyecare practitioner should provide the patient with the wearer's guide that pertains to the patients prescribed lens.

DESCRIPTION:

The **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft (hydrophilic) Contact Lens** is available as a single vision spherical lens, and as a back surface astigmatic (toric) lens. The hydrophilic nature of this material allows the lens to become soft and pliable when immersed in an aqueous solution.

The non-ionic lens material, (hioxifilcon A) is a random copolymer of 2-hydroxyethyl methacrylate and 2,3-dihydroxypropyl methacrylate, cross linked with ethylene glycol dimethacrylate. It consists of 43% hioxifilcon A and 57% water by weight when immersed in buffered saline solution. The lens is available clear or with a blue visibility-handling tint, color additive 'Reactive Blue 19', 21 CFR part 73.3121. The United States Adopted Names Council (USAN) has adopted the (hioxifilcon A) name.

In the hydrated state, the lens conforms to the curvature of the eye covering the cornea and extending slightly beyond the limbus forming a colorless, transparent optical surface. The hydrophilic properties of the lens require that it be maintained in a fully hydrated state in a solution compatible with the eye. If the lens dries out, it will become hard and appear somewhat warped however, it will return to its proper configuration when completely rehydrated in the proper storage solution.

The hydrophilic characteristics allow aqueous solutions to enter the lens and in its fully hydrated state the lens is approximately 57% water by weight. The physical properties of the lens are:

Refractive Index	1.4058 (wet)
Light Transmission	greater than 95%
Water Content	57 %
Specific Gravity	1.1287 (hydrated)
Oxygen Permeability	20.0 x 10 ⁻¹¹ (cm ² /sec) (ml O ₂ /ml x mm Hg @ 35°C), (Fatt method ISO9913-1).

The lenses are hemispherical flexible shells which cover the cornea and portion of the adjacent sclera with the following dimensions:

Chord Diameter	: 14.2-14.6 mm
Center Thickness	: 0.08 mm
Base Curve	: 8.5-8.9mm
Powers	: 0.0 to -6.00 Diopters in 0.25 D increments -6.50 to -20.00 Diopters in 0.50 D increments

	+0.25 to +4.00 Diopters in 0.25 D increments
	+4.50 to +20.00 Diopters in 0.50 D increments
Cylinders	: -1.00 D, -1.75 D, -2.50 D
Axis	: 80°, 90°, 100°, 20°, 10°, 180°, 170°, 160°

ACTIONS:

In its hydrated state, the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens**, when placed on the cornea, act as a refracting medium to focus light rays on the retina. The (toric) lens provides a more even surface over the highly uneven astigmatic cornea and thus helps to focus light rays on the retina.

INDICATIONS (USES):

The **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** for daily wear are indicated for the correction of visual acuity in aphakic and not aphakic persons with non-diseased eyes with myopia or hyperopia and may have astigmatism of 7.00D of astigmatism or less. The lens may be cleaned and disinfected using a chemical (not heat) lens care system.

Eyecare practitioners may prescribe the lenses for frequent/planned replacement wear, with cleaning, disinfection and scheduled replacement (see WEARING SCHEDULE). When prescribed for frequent/planned replacement wear, the lens may be cleaned and disinfected using a chemical (not heat) lens care system.

CONTRAINDICATIONS (REASONS NOT TO USE):

DO NOT USE the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** when any of the following conditions exist:

- Acute and subacute inflammation or infection of the anterior chamber of the eye.
- Any eye disease, injury, or abnormality that affects the cornea, conjunctiva, or eyelids.
- Severe insufficiency of lacrimal secretion (dry eyes).
- Corneal hypoesthesia (reduced corneal sensitivity), if not-aphakic.
- Any systemic disease that may affect the eye or be exaggerated by wearing contact lens.
- Allergic reactions of ocular surfaces or adnexa that may be induced or exaggerated by wearing contact lens or use of contact lens solutions.
- Allergy to any ingredient, such as mercury or thimerosal, in a solution which is to be used to care for the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens**.
- Any active corneal infection (bacterial, fungi, or viral)
- If eyes become red or irritated.
- Patients unable to follow lens care regimen or unable to obtain assistance to do so.

WARNINGS:

- PROBLEMS WITH CONTACT LENS AND LENS CARE PRODUCTS COULD RESULT IN **SERIOUS INJURY TO THE EYE**. It is essential that patients follow their eyecare practitioner's direction and all labeling instructions for proper use of lens and lens care products, including the lens case. **EYE PROBLEMS, INCLUDING CORNEAL ULCERS, CAN DEVELOP RAPIDLY AND LEAD TO LOSS OF VISION.**
- All contact lens wearers must see their eyecare practitioner as directed.
- Daily wear lens are not indicated for overnight wear, and patients should be instructed not to wear lens while sleeping. Clinical studies have shown that the risk of serious adverse reactions is increased when this lens is worn overnight.
- Studies have shown that contact lens wearers who are smokers have a higher incidence of adverse reactions than nonsmokers do.
- **IF YOU EXPERIENCE EYE DISCOMFORT, EXCESSIVE TEARING, VISION CHANGES, OR REDNESS OF THE EYE, IMMEDIATELY REMOVE YOUR LENS AND PROMPTLY CONTACT YOUR EYECARE PRACTITIONER.**

PRECAUTIONS:

Special Precautions for eyecare practitioner:

- Clinical studies have demonstrated that contact lens manufactured from (hioxifilcon A) is safe and effective for their intended use. However, the clinical studies may not have included all design configurations or lens parameters that are presently available in this lens material.

Due to the small number of patients enrolled in clinical investigation of lenses, all refractive powers, design configurations, or lens parameters available in the lens materials are not evaluated in significant numbers. Consequently, when selecting an appropriate lens design and parameters, the eyecare practitioner should consider all characteristics of the lens that can affect lens performance and ocular health, including oxygen permeability, wettability, central and peripheral thickness, and optic zone diameter.

The potential impact of these factors on the patient's ocular health should be carefully weighed against the patient's need for refractive correction: therefore, the continuing ocular health of the patient and lens performance on the eye should be carefully monitored by the prescribing eyecare practitioner.

- Aphakic patients should not be fitted with **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** until the determination is made that the eye has healed completely.
- Fluorescein, a yellow dye, should not be used while the lens is on the eye. The lens absorbs this dye and become discolored. Whenever Fluorescein is used in eyes, the eyes should be flushed with a sterile saline solution that is recommended for in eye use. Wait at least one hour before replacing the lens. Too early replacement may allow the lens to absorb residual Fluorescein irreversibly.
- Before leaving the eyecare practitioner's office, the patient should be able to promptly remove lens or should have someone else available who can remove the lens for him or her.
- Eyecare practitioners should instruct the patient to remove the lens immediately if the eye becomes red or irritated.

Eyecare practitioners should carefully instruct patients about the following care regimen and safety precautions:

- Different solutions cannot always be used together, and not all solutions are safe for use with all lens. Use only recommended solutions that are fresh and sterile.
 - Never use solutions recommended for conventional hard contact lens only.
 - Chemical disinfection solutions should not be used with heat unless specifically indicated on product labeling for use in both heat and chemical disinfection.
 - Always use **FRESH, STERILE UNEXPIRED** lens care solutions.
 - Always follow directions in the package inserts for the use of contact lens solutions.
 - Use only a chemical (not heat) lens care system. Use of the heat (thermal) care system can discolor the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens**.
 - Sterile unpreserved solutions, when used, should be discarded after the time specified in the labeling directions.
 - Do not use saliva or anything other than the recommended solution for lubricating or rewetting lens.
 - Always keep the lens completely immersed in the recommended storage solution when the lens is not being worn (stored). Prolonged periods of drying will damage the lens. Follow the lens care directions for (Care for a Dried Out Dehydrated Dry Lens) if the lens surface does become dried out.
- If the lens sticks (stops moving) on the eye, follow the recommended directions on (care for sticking non-moving lens). The lens should move freely on the eye for the continued health of the eye. If nonmovement of the lens continues, the patient should be instructed to **IMMEDIATELY** consult his or her eyecare practitioner.

- Always wash and rinse hands before handling lens. Do not get cosmetics, lotions, soaps, creams, deodorants, or sprays in the eyes or on the lens. It is best to put on lens before putting on makeup. Water-base cosmetics are less likely to damage lens than oil-base products.
- Do not touch contact lens with the fingers or hands if the hands are not free of foreign materials, as microscope scratches of the lens may occur, causing distorted vision and/or injury to the eye.
- Carefully follow the handling, insertion, removal, cleaning, disinfection, storing and wearing instructions in the patient instructions for the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** as those prescribed by the eyecare practitioner.
- Never wear lens beyond the period recommended by the eyecare practitioner.
- If aerosol products such as hair spray are used while wearing lens, exercise caution and keep eyes closed until the spray has settled.
- Always handle lens carefully and avoid dropping them.
- Avoid all harmful or irritating vapors and fumes while wearing lens.
- Ask the eyecare practitioner about wearing lens during sporting activities.
- Inform the doctor (health care practitioner) about being a contact lens wearer.
- Never use tweezers or other tools to remove lens from the lens container unless specifically indicated for that use. Pour the lens into the hand.
- Do not touch the lens with fingernails.
- Always discard disposable lens and lens worn on a frequent replacement schedule after the recommended wearing schedule prescribed by the eyecare practitioner.
- Always contact the eyecare practitioner before using any medicine or medications in the eyes.
- Always inform the employer of being a contact lens wearer. Some jobs may require use of eye protection equipment or may require that the patient not wear contact lens.
- As with any contact lens, follow-up visits are necessary to assure the continuing health of the patient's eyes. The patient should be instructed as to a recommended follow-up schedule.

ADVERSE REACTIONS:

The patient should be informed that the following problems may occur:

- Eyes stinging, burning, itching (irritation), or other eye pain.
- Comfort is less than when lens was first placed on eye.
- Feeling that something is in the eye such as a foreign body or scratched area.
- Excessive watering (tearing) of the eye.
- Unusual eye secretions.
- Redness of the eye.
- Reduced sharpness of vision (poor visual acuity).
- Blurred vision, rainbows, or halos around objects.
- Sensitivity to light (photophobia).
- Dry eyes.

If the patient notices any of the above, he or she should be instructed to:

- **IMMEDIATELY REMOVE LENS.**
- If discomfort or problems stops, then look closely at the lens. If the lens is in any damage, **DO NOT PUT THE LENS BACK ON THE EYE.** Place the lens in the storage case and contact the eyecare practitioner. If the lens has dirt, an eyelash, or other foreign body on it, or the problem stops and the lens appears undamaged, the patient should thoroughly clean, rinse, and disinfect the lens then reinsert them. After reinsertion, if the problem continues, the patient should **IMMEDIATELY REMOVE THE LENS AND CONSULT THE EYECARE PRACTITIONER**

When any of the above problems occur, a serious condition such as infection, corneal ulcer, neovascularization, or iritis may be present. The patient should be instructed to **KEEP LENS OFF THE EYE AND SEEK IMMEDIATE PROFESSIONAL IDENTIFICATION** of the problem and prompt treatment to avoid serious eye damage.

FITTING:

Conventional methods of fitting contact lens apply to the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens**. For a detailed description of the fitting techniques, refer **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens Professional Fitting Guide and Patient Information/ Wearer's Guide**, copies of which are available from:

Clearlab US Inc.

4200 Jenkins Court, Suwanee,
GA 30024, United States of America
Tel: +1 770 2710211
Fax: +1 770 2710225
Email: USRA@clearlab.com
Website: www.clearlabus.com

RECCOMENDED WEARING SCHEDULE:

Close professional supervision is recommended to ensure safe and successful contact lens wear. If the patient complains of discomfort, decreased vision, ocular injection or corneal edema, the lens should be removed and the patient scheduled for examination. The problem may be relieved by putting the patient on a different wearing schedule or possibly by refitting the lens.

THE WEARING AND REPLACEMENT SCHEDULES SHOULD BE DETERMINED BY THE EYECARE PRACTITIONER.

Patients tend to overwear the lens initially. The eyecare practitioner should emphasize the importance of adhering to the initial maximum wearing schedule. It is important not to

exceed the wearing schedule. Regular check ups, as determined by the eyecare practitioner, are also extremely important.

The maximum suggested wearing schedule for the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** is suggested below.

DAY	1	2	3	4	5	6
HOURS	6	8	10	12	14	All waking hours

STUDIES HAVE NOT BEEN COMPLETED TO SHOW THAT THE “eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens” IS SAFE TO WEAR DURING SLEEP.

FREQUENT / PLANNED REPLACEMENT:

It is recommended that the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** be discarded and replaced with a new lens every two weeks. However, as the eyecare practitioner, you are encouraged to determine an appropriate lens replacement schedule based upon the response of the patient.

LENS CARE DIRECTIONS:

Eyecare practitioners should review with the patient lens care directions, including both basic lens care information and specific instructions on the lens care regimen recommended for the patient:

General Lens Care (To First Clean and Rinse, Then Disinfect Lenses)

Basic Instructions:

- Care of contact lens takes very little time and involves **THREE** essential steps - **CLEANING, RINSING AND DISINFECTING**. Each step in itself is important, and one step is not to be replaced by the other.
- Always wash, rinse and dry hands before handling contact lens.
- Always use **FRESH, STERILE UNEXPIRED** lens care solutions.
- Use the recommended lens care system; either chemical (not heat) or heat (thermal). Different solutions cannot always be used together, and not all solutions are safe for use with all lens. **DO NOT ALTERNATE OR MIX LENS CARE SYSTEMS UNLESS INDICATED ON SOLUTION LABELING.**
- Do not use saliva or anything other than the recommended solutions for lubricating or rewetting lens. Do not put lens in the mouth.
- Lens should be **cleaned, rinsed, and disinfected** each time they are removed. **Cleaning and rinsing** are necessary to remove mucus and film from the lens surface. **Disinfecting** is necessary to destroy harmful germs.
- Always remove, clean, rinse, enzyme (as recommended by the eyecare practitioner) and disinfect lenses according to the schedule prescribed by the

eyecare practitioner. The use of an enzyme or any cleaning solution does not substitute for disinfection.

- The eyecare practitioner should recommend a care system that is appropriate for **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens**. Each lens care product contains specific directions for use and important safety information, which should be read and carefully followed.

Note: Some solutions may have more than one function, which will be indicated on the label. Read the label on the solution bottle, and follow instructions.

- **Lens cleaning, disinfection, and storage:**

- **Clean** one lens first (always the same lens first to avoid mix-ups), rinse the lens thoroughly with recommended rinsing or disinfecting solution to remove the cleaning solution, mucus, and film from the lens surface, and put lens into correct chamber of the lens storage case. Then repeat the procedure for the second lens.
- After cleaning, **disinfect** lens using the system recommended by the manufacturer and/or the eyecare practitioner.
- To store lens, disinfect and leave them in the closed/unopened case until ready to wear. If lens is not to be used immediately following disinfection, the patient should be instructed to consult the package insert or the eyecare practitioner for information on storage of lens. Follow the instruction and timing recommended by the solution manufacturer.
- After removing the lens from the lens case, empty and rinse the lens storage case with solution as recommended by the lens case manufacturer; then allow the lens case to air dry. When the case is used again, refill it with storage solution as recommended by the lens case manufacturer; or your eyecare practitioner.
- Eyecare practitioners may recommend a lubricating/rewetting solution which can be used to wet (lubricate) lens while they are being worn to make them more comfortable.

- **Chemical (NOT HEAT) Lens Disinfection:**

1. Wash and rinse your hands thoroughly BEFORE HANDLING LENS.
2. After removal of lens, **CLEAN** the lens by applying three drops of cleaner to each surface. Then rub the lens between your fingers for 20 seconds.
3. **AFTER CLEANING**, thoroughly rinse both surfaces of the lens with a steady stream of fresh, sterile rinsing solution for approximately 10 seconds.
4. Fill contact lens carrying case with the recommended disinfection and storage solution and place lens in the proper cells for a minimum of 4 hours. Follow the instruction and timings recommended by the manufacturer or eyecare practitioner.
5. When use hydrogen peroxide lens care systems, lenses must be neutralized before wearing. Follow the recommendations on the hydrogen peroxide system labeling.

6. Thoroughly rinse lens with a fresh solution recommended for rinsing before inserting and wearing, or follow the instructions on the disinfection solution labeling.
7. Leave the lens in the unopened storage case until ready to put on the eyes.

Note: **DO NOT HEAT THE DISINFECTION SOLUTION AND LENS.**

Caution: Lenses that are chemically disinfected may absorb ingredients from the disinfecting solution, which may be irritating to the eyes. A thorough rinse in fresh, sterile rinsing solution prior to placement on the eye should reduce the potential for irritation.

- **Lens Care Regimen:**

Patients must adhere to the lens care regimen recommended by their eyecare practitioner for the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens**. Failure to follow this procedure may result in development of serious ocular infections

- **Storage:**

The **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** must be stored only in the recommended solutions. If left exposed to the air, the lens will dehydrate. If lens dehydrates, reference above section on caring for dried out (dehydrated) dry lens.

- **Lenses prescribed for frequent replacement:**

The **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** may be prescribed in a frequent replacement program and should be thrown away after the recommended wearing period prescribed by the eyecare practitioner.

LENS DEPOSITS AND USE OF ENZYMATIC CLEANER:

Enzyme cleaning may be recommended by the eyecare practitioner. Enzyme cleaning removes protein deposits on the lens. These deposits cannot be removed with regular cleaners. Removing protein deposits is important for the well being of the patient's lens and eyes. If these deposits are not removed, they can damage the lens and cause irritation.

Enzyme cleaning does NOT replace routine daily cleaning and disinfecting. For enzyme cleaning, the patient should carefully follow the instructions in the enzymatic cleaning labeling.

LENS CASE CLEANING AND MAINTENANCE:

Contact lens cases can be a source of bacteria growth. Lens cases should be emptied, cleaned, rinsed with solutions recommended by the lens case manufacturer, and allowed to air dry. When the case is used again, refill it with storage solution. Lens case should be replaced at regular intervals as recommended by the lens case manufacturer or your eyecare practitioner.

CARE FOR A DRIED OUT (DEHYDRATED) DRY LENS:

If for some reason, your lens dries out completely a minimum of handling is important, as they are very brittle in the dehydrated state. Carefully place them in rinsing or storage solution for a minimum of thirty minutes during which time they will become soft and flexible. Then follow the cleaning, rinsing, and disinfecting procedures - including soaking the lens in storage and disinfection solution for four hours before wearing again.

CARE FOR A STICKING (NONOVING) LENS:

If the lens sticks (cannot be removed), the patient should be instructed to apply 3 to 4 drops of the recommended lubricating or rewetting solution directly to the eye and wait until the lens begins to move freely on the eye before removing it. If non-movement of the lens continues after 15 minutes, the patient should **IMMEDIATELY** consult the eyecare practitioner.

RECOMMENDED LENS CARE PRODUCTS:

The eyecare practitioner should recommend a care system that is appropriate for the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens**. Each lens care product contains specific instructions for use and important safety information, which should be read and carefully followed.

EMERGENCIES:

The patient should be informed that if chemicals of any kind (household products, gardening solutions, laboratory chemicals, etc.) are splashed into the eyes, the patient should:

FLUSH EYES IMMEDIATELY WITH TAP WATER AND IMMEDIATELY CONTACT THE EYECARE PRACTITIONER OR VISIT A HOSPITAL EMERGENCY ROOM WITHOUT DELAY.

HOW SUPPLIED:

Each lens is supplied sterile in blister packs in bicarbonate buffered aqueous solution with poloxamer. The blister pack is marked with the base curve, diameter, dioptric power, manufacturing lot number, and expiration date of the lens.

REPORTING OF ADVERSE REACTIONS:



All serious adverse experiences and adverse reactions observed in patients wearing the **eyedia® precise monthly for astigmatism (Hioxifilcon A) Soft Contact Lens** or experienced with the lens should be reported to:

Clearlab US Inc.

4200 Jenkins Court, Suwanee,
 GA 30024, United States of America
 Tel: +1 770 2710211
 Fax: +1 770 2710225
 Email: USRA@clearlab.com
 Website: www.clearlabus.com

SYMBOLS KEY:

The following symbols may appear on the label or carton.

SYMBOL	DESCRIPTION
	Sterile Using Steam
LOT	Product Lot Number
EXP	Expiry Date
BC	Lens Base Curve
DIA	Lens Diameter
CYL	Lens Cylinder
AXIS	Lens Axis
Rx Only	Caution: Federal law restricts this device to sale by or on the order of a licensed Eye Care Practitioner.
 SINGLE PATIENT USE	Caution: This is a single patient use device; See Package Insert or Instructions For Use.

CAUTION: Federal law restricts this device to sale by or on the order of a licensed practitioner.



Clearlab US Inc.
 Printed in Singapore
 Revision Date: 18 October 2011
 Doc Number: S-ASP-023-A
 Version Number: V02

© 2011 CLEARLAB SG PTE. LTD. All Rights Reserved.